桂林市啄木鸟医疗器械有限公司
GUILIN WOODPECKER Medical Instrument Co.,LTD.

File No.: ZMN-YF-(D01)-01-001-02
Version: B

# EC Declaration of Conformity

*Manufacturer:*

**Guilin Woodpecker Medical Instrument Co., Ltd.**

*whose single Authorized Representative:*

**MedNet EC-Rep GmbH • Borkstrasse 10 • 48163 Muenster • Germany**

We, the manufacturer, herewith declare that the products
**Dental Implant Unit, UMDNS-***Code:* 23653

| Product name | Model |
|---|---|
| Dental Implant Unit | Implanter、Implanter LED |
| | Implant-X、Implant-X LED |

meet the provisions of Directive 93/42/EEC which apply to them.

The medical device has been assigned to class IIa according to Annex IX of the Directive 93/42/EEC. It bears the mark

CE 0197

The product concerned has been manufactured under a quality management system according to Annex V and Annex VII of Directive 93/42/EEC.

Compliance of the designated product with the Directive 93/42/EEC has been assessed and certified by the Notified Body

**TÜV Rheinland LGA Products GmbH**
**Tillystraße 2, 90431, Nürnberg, Germany**
Certificate No.: DD60137494 0001
Issue date: 2019-07-16
Expiry date: 2024-05-27

following the procedure relating to the EC Declaration of Conformity set out in Annex V and Annex VII of Directive 93/42/EEC.

This Declaration of Conformity covers all medical devices as specified in the product list belonging to this declaration and is only valid in connection with a batch specific Certificate of Compliance for all products concerned bearing the CE mark

The above mentioned declaration of conformity is exclusively under the responsibility of

Company: **Guilin Woodpecker Medical Instrument Co., Ltd.**
Address: Information Industrial Park, GuiLin National High-Tech Zone, GuiLin, GuangXi, 541004, P.R.China

杨芸凤 2020.3.19
Preparation, *date*

2020.3.19
Review, *date*

2020.3.19
*Legally binding signature, Function*